

Methodology of ^1H NMR Spectroscopy of the Human Brain at Very High Magnetic Fields

I. Tkáč and R. Gruetter

Center for Magnetic Resonance Research, University of Minnesota, Minneapolis, Minnesota, USA

Received August 4, 2004; revised September 30, 2004

Abstract. An ultrashort-echo-time stimulated echo-acquisition mode (STEAM) pulse sequence with interleaved outer volume suppression and VAPOR (variable power and optimized relaxation delays) water suppression was redesigned and optimized for human applications at 4 and 7 T, taking into account the specific requirements for spectroscopy at high magnetic fields and limitations of currently available hardware. In combination with automatic shimming, automated parameter adjustments and data processing, this method provided a user-friendly tool for routine ^1H nuclear magnetic resonance (NMR) spectroscopy of the human brain at very high magnetic fields. Effects of first- and second-order shimming, single-scan averaging, frequency and phase corrections, and eddy currents were described. LCModel analysis of an in vivo ^1H NMR spectrum measured from the human brain at 7 T allowed reliable quantification of more than fifteen metabolites noninvasively, illustrating the potential of high-field NMR spectroscopy. Examples of spectroscopic studies performed at 4 and 7 T demonstrated the high reproducibility of acquired spectra quality.

1 Introduction

The first pioneers in nuclear magnetic resonance (NMR) spectroscopy of liquids discovered a long time ago that an increased magnetic field improved sensitivity and spectral resolution. Fifty years of technological and methodological development has resulted in NMR spectrometers operating at 900 MHz (21 T) used in biochemistry to elucidate the structure of proteins and nucleic acids in solutions [1]. In vivo NMR spectroscopy requires wide-bore-size magnets equipped with a field gradient system for spatial localization. Such instrumentation is typical for magnetic resonance imaging (MRI) scanners. MRI has developed into a standard clinical imaging procedure [2], but the majority of currently installed MRI scanners for humans operate at magnetic fields below 3 T. Recent progress in magnet technology, gradient system performance, and radio-frequency (RF) coil design enabled localized in vivo NMR spectroscopy of humans at magnetic fields up to 7 T [3–6]. However, standard localization techniques, such as point-resolved spectroscopy (PRESS) and stimulated-echo acquisition mode (STEAM) [7–9],

based on band-selective RF pulses and gradients were developed using "low-field" MR scanners. Pulse sequences designed for single-voxel NMR spectroscopy at high magnetic fields must take into account increased off-resonance effects at high fields that cause chemical shift displacement errors. In addition, high-field NMR spectroscopy of humans has specific requirements for the hardware (performance of the gradient system and RF coils), elimination of B_0 inhomogeneity (shimming), specific absorption rate (SAR), etc.

The aim of this paper is to provide a detailed description of a pulse sequence specifically designed for localized ^1H NMR spectroscopy of humans at very high magnetic fields. An ultrashort-echo-time STEAM sequence combined with outer volume suppression (OVS) and VAPOR (variable power and optimized relaxation delays) water suppression, which originally designed for spectroscopy of rats [10] and mice [11] at 9.4 T, was adapted for human applications at 4 and 7 T. Initial results have been shown several years ago [12–15] and since then this sequence has been routinely used [13, 16, 17]. In this article we summarize our experience with high-field ^1H NMR spectroscopy of humans and highlight the most important features of pulse sequence design, data acquisition, data processing, metabolite quantification, and present some examples of applications at 4 and 7 T.

2 Methods

2.1 Instrumentation

The measurements described herein were performed on 4 and 7 T whole-body systems. The 4 T, 90 cm magnet (Siemens/Oxford Magnet Technology, Oxford, UK) was equipped with a Sonata gradient system (40 mT/m, 400 μs rise-time; Siemens AG, Erlangen, Germany), which included second-order shim coils with maximum shim strengths of $XZ = 1.1 \cdot 10^{-4}$ mT/cm², $YZ = 1.1 \cdot 10^{-4}$ mT/cm², $Z^2 = 2.8 \cdot 10^{-4}$ mT/cm², $2XY = 5.8 \cdot 10^{-5}$ mT/cm², $X^2 - Y^2 = 5.3 \cdot 10^{-5}$ mT/cm² at a current of 20 A for Z^2 coil and 10 A for all other second-order shim coils. The 7 T, 90 cm magnet (Magnex Scientific, Yarnton, UK) was equipped with a head gradient coil (40 mT/m, 500 μs rise-time) driven by a Harmony/Symphony gradient amplifier (Siemens AG, Erlangen, Germany), and strong-custom designed second-order shim coils (Magnex Scientific, UK) with maximum strengths of $XZ = 5.8 \cdot 10^{-4}$ mT/cm², $YZ = 5.6 \cdot 10^{-4}$ mT/cm², $Z^2 = 9.0 \cdot 10^{-4}$ mT/cm², $2XY = 2.8 \cdot 10^{-4}$ mT/cm², $X^2 - Y^2 = 2.9 \cdot 10^{-4}$ mT/cm² at a current of 4 A. Shim coil currents were controlled with amplifiers from Resonance Research, Inc (Billerica, MA). All first- and second-order shim terms were automatically adjusted by FASTMAP with EPI readout [18, 19]. Both magnets were interfaced to a Varian INOVA console (Varian Inc., Palo Alto, CA). The RF amplifiers (CPC, Brentwood, NY) provided maximum power of 2.6 kW (4 T) and 4.5 kW (7 T) at the coil. The highest achievable $\gamma B_1/2\pi$ at 4 and 7 T was approximately 2 kHz using quadrature transmit-receive half-volume RF coils

with two geometrically decoupled single-turn coils with a diameter of 12 cm [20]. For the measurement of ^1H NMR spectra from deeper brain regions at 4 T, a TEM coil [21] was used (maximum $\gamma B_1/2\pi$ of about 1 kHz). The position of the volume of interest (VOI) was based on multislice RARE (rapid acquisition relaxation-enhanced) imaging with the following parameters: ETL (echo train length), 8; echo spacing, 15 ms; TE (echo time), 60 ms; number of phase-encoding steps, 256; FOV (field of view), 20 by 20 cm; slice thickness, 4 mm; NT (number of transients), 2; TR (repetition time), 4 s. Specific absorption rates (SAR) were maintained within u.s. Food and Drug Administration guidelines for all experiments.

2.2 Quantification of Metabolites

In vivo ^1H NMR spectra were analyzed using LCMModel [22–24]. The unsuppressed water signal measured from the same VOI was used as an internal reference for quantification assuming a brain water content of 80%. The LCMModel basis set included spectra of 20 brain metabolites: alanine, ascorbate (Asc), aspartate (Asp), creatine (Cr), phosphocreatine (PCr), γ -aminobutyric acid (GABA), glucose (Glc), glutamate (Glu), glutamine (Gln), glutathione (GSH), glycerophosphorylcholine (GPC), phosphorylcholine (PCho), glycine, *myo*-inositol (Ins), lactate (Lac), N-acetylaspartate (NAA), N-acetylaspartylglutamate (NAAG), phosphorylethanolamine (PE), *scyllo*-inositol, and taurine (Tau). Spectra were simulated with Varian spectrometer spin-simulation software (neglecting J-evolution for ultrashort TE) based on previously published chemical shifts and coupling constants [25]. Chemical shifts of methylene protons of PCho were corrected using our experimental phantom data [24] and missing geminal coupling constants of GABA, GPC, PCho, PE, and Tau were set to -15 Hz. All signals within the chemical shift range of 4.68 ± 0.35 ppm were eliminated from simulated spectra due to frequency selectivity of VAPOR water suppression pulses. A spectrum of fast-relaxing macromolecules experimentally measured from the human brain by an inversion-recovery experiment (TR = 2 s, TI(inversion time) = 0.675 s) was also included in the basis set.

2.3 In Vivo Studies

All subjects were studied after giving informed consent according to procedures approved by Institutional Review Board at the University of Minnesota. Typical voxel sizes were 15 ml at 4 T and 8 ml at 7 T. TR was 4.5 s at 4 T and 5.0 s at 7 T and the echo time of STEAM was between 4 and 6 ms (see also Sect. 3).

3 Results

3.1 Pulse Sequence Design

In order to minimize the chemical shift displacement error (off-resonance effects on spatial localization) within the limited peak B_1 and to maintain the selectivity of water suppression (± 0.35 ppm), the pulse sequence previously developed for the 9.4 T animal MR scanner [10] had to be redesigned as shown in the pulse sequence scheme (Fig. 1). The VAPOR water suppression pulse train [10] was interleaved with 4 blocks of outer volume suppression (OVS) and followed by ultrashort echo time STEAM for localization.

3.2 Ultrashort-Echo-Time STEAM

As in the previous localization method, asymmetric RF pulses were used to reduce the TE [10]. To allow an optimal power adjustment for slightly different loading conditions of the RF coil, the pulse width was set to a value that required the peak RF power (for a 90° pulse) to be a few decibel less than the maximum RF power available. The default duration of the asymmetric STEAM pulses was set to 1.5 ms at 7 T, which required a peak $\gamma B_1/2\pi = 1.12$ kHz and resulted in 4.5 kHz bandwidth. The pulse width for the 4 T system was set to 2.0 ms (peak $\gamma B_1/2\pi = 0.84$ kHz, 3.4 kHz bandwidth), in order to make the pulse sequence usable for the TEM volume coil. Bandwidths of STEAM pulses (at 50% of the excitation profile maximum) were used to set the slice-selection gradient




Fig. 1. Schematic illustration of the localization pulse sequence designed for ultrashort echo time in vivo ^1H NMR spectroscopy of human brain at high magnetic fields (4–7 T). The VAPOR water suppression is interleaved with OVS. The STEAM sequence was designed to achieve a minimum echo time of 4–6 ms.

amplitudes for requested sizes of the VOI. The minimum TE was between 4 and 6 ms, depending on the duration of crusher gradients required for the elimination of unwanted coherences. In general, requirements for the dephasing power of crusher gradients (amplitude, duration) increased with the sensitive volume of the RF coil and B_0 . A soft Gaussian RF pulse was used during the TM period for the ultimate suppression of the residual water signal remaining after VAPOR water suppression. To keep the frequency selectivity at different field strengths approximately the same (± 0.35 ppm), the duration of this Gaussian pulse was set to 30 ms (4 T) and 20 ms (7 T), which resulted in a TM period duration of 42 ms (4 T) and 32 ms (7 T). Data acquisition started before the maximum of the stimulated echo (STE) to avoid possible baseline distortions due to incorrect amplitudes of the first data points of the free induction decay (FID) (immediately after turning on the receiver). Data points acquired before the maximum of the STE were discarded in postprocessing. The timing of the first data point of the STE was carefully adjusted to eliminate the need for any first-order phase correction in the resulting spectra. RF power for 90° STEAM pulses was calibrated for each VOI as follows: RF power was incremented by 1 dB and the value corresponding to the maximum signal intensity was automatically selected. This power was the basis for setting the power for OVS and VAPOR pulses.

3.3 Outer Volume Suppression

OVS was necessary to improve the localization performance, i.e., elimination of signals arising from outside of the VOI. The OVS scheme with three OVS blocks, originally developed for animal studies at 9.4 T [10], was redesigned for human application. The last OVS block that saturated 6 slabs around the VOI [10] was substituted by two OVS blocks, each of them saturated 4 slabs (Fig. 1). This modification was made due to the timing requirements of VAPOR and the duration of RF pulses in the OVS pulse train. Hyperbolic-secant RF pulses with the R-value of 42 (the product of band-width and pulse width and a nominal flip angle of about 90° were used for the OVS pulses. The 8.4 kHz bandwidth of OVS RF pulses (pulse duration of 5.0 ms, RF power ranging from -5 dB to +1 dB relative to the peak power of the STEAM 90° pulse) reduced the chemical shift displacement error of the OVS to values comparable with chemical displacement error of the VOI. The RF power applied to the second OVS pulse train was increased by 3 dB relative to the first one to improve the OVS efficiency in the presence of a spatially inhomogeneous B_1 , frequently encountered at high magnetic fields. Crusher gradients were carefully set to avoid unwanted coherence rephasing. The optimal RF power for the OVS pulses was measured experimentally with profiles along X , Y , and Z axes centered on the VOI. Once optimal settings were found, values of the RF power for each individual OVS pulse were fixed relative to the power of the 90° STEAM pulse. This approach guaranteed that optimal values of the RF power were always used for OVS,

independent of the loading of the RF coil. The localization performance of the sequence (STEAM with OVS) was extensively tested using two-chamber phantoms (10 mM Cr in an inner chamber, 25% ethanol in water or vegetable oil in an outer chamber).

3.4 VAPOR Water Suppression

The limited RF power and larger sensitive volume of RF coils on human MR systems resulted in lower peak $\gamma B_1/2\pi$ available that required longer RF pulses in OVS. An increased duration of the OVS pulse train consequently required changes in original VAPOR inter-pulse timing scheme [10]. In addition, to maintain the same selectivity for water suppression (± 0.35 ppm) as in the original implementation at 9.4 T, lower magnetic field required longer RF pulses. The key principle of the VAPOR water suppression is to set optimal power variation and time delays in a multipulse train. Taking into account the timing limitations, a simulation of the VAPOR water suppression efficiency was performed to find the optimal combination of RF power variations and the timing to minimize the sensitivity of VAPOR on T_1 and B_1 . Truncated (3 lobes) asymmetric RF pulses [10] of 35 ms (4 T) and 25 ms (7 T) duration were used in the VAPOR water suppression, which represented a reasonable compromise between the excitation profile of the pulse and the timing requirement of the interleaved water suppression and OVS pulse train. Optimized parameters of the VAPOR water suppression for 4 and 7 T are shown in Table 1. The ratio between the power for the STEAM 90° pulse and the optimal power for the nominal VAPOR pulse (N in Table 1) was calculated and experimentally verified. This ratio was independent of RF coil loading, consistent with theory. Therefore, the optimal RF power for VAPOR water suppression was automatically set when the water suppression was turned on. Fine adjustment of water suppression was possible by varying the delay between the 7th and the 8th pulse of VAPOR, which was not necessary for most applications. The crusher gradients in VAPOR were adjusted to minimize the chance for rephasing of unwanted coherences (mainly stimulated echoes) of the water resonance. Using different crusher gradient orientations and

Table 1. RF power and inter-pulse delays in VAPOR water suppression pulse train with 8 RF pulses optimized for 4 and 7 T.

Parameter	Pulse/delay number							
	1	2	3	4	5	6	7	8
RF power (dB) ^a , 4 T, 7 T	N	N	N + 5	N	N + 4	N	N + 5	N + 5
Interpulse delay (ms)								
4 T	150	100	140	105	105	66	72	18
7 T	150	100	122	105	102	61	67	14

^a N is the nominal value of the peak RF power of frequency-selective water suppression pulses.

amplitudes minimized the probability of unwanted echoes. Settings of the crusher gradients were verified experimentally by incrementing the amplitude of all three crusher gradients (along X , Y , and Z) between the last RF pulse of VAPOR and the first RF pulse of STEAM and by searching for possible echoes of the water resonance. Amplitudes of these crusher gradients were then set to values that were as far away as possible from values generating echoes.

3.5 Single-Scan Data Acquisition

Physiological motion, such as small random motions of the head or motion of the diaphragm during breathing, resulted in frequency and phase variation in spectra (Fig. 2). The impact of frequency and phase fluctuations on spectral quality (reduced signal-to-noise ratio [SNR], increased line width) was easily minimized (Fig. 2) by saving each acquired FID separately on disk and by post-processing corrections before summation. Frequency and phase corrections were possible provided that SNR was ≥ 3 in single-scan spectra. The frequency variation was calculated from the signal of NAA in absolute-value mode; however, the frequency and phase corrections were performed in the time domain. This approach was used for all spectroscopy studies at 7 T and some studies at 4 T. Single-scan spectra were routinely very reproducible and free from artifacts, e.g., contamination by signal from outside of the VOI (subcutaneous lipids) and/or insufficiently suppressed water signal coherences (Fig. 3). The residual water signal was always substantially lower than the signal of NAA, even though the only experimentally adjusted parameter was the RF power of the STEAM 90° pulse. In addition, a 32-step phase cycle was implemented to subtract potential unwanted coherences from outside of the VOI.

3.6 Spectral Quality

Data processing included automatic frequency and phase correction of the single-scan data, summation of FIDs, elimination of residual eddy current effects using the water signal from the corresponding VOI [26], and automatic zero-order phase correction (subtraction of the phase of the first point from all points of FID). Neither water signal removal nor baseline correction was applied. The quality of in vivo ^1H NMR spectra from the human brain currently achievable on our 7 T magnet with the described localization sequence is illustrated in Fig. 4. The position and size of the VOI are shown on sagittal and transverse images of the brain. Characteristic signals of NNA, Cr, cholines (Cho), Glu, Ins, and Tau are clearly discernible (Fig. 4). In addition, the H-1 signal of α -D-glucose was observable, demonstrating the efficiency and selectivity of the VAPOR water suppression.

Turning off the outer volume suppression resulted in contamination of spectra by signals of subcutaneous lipids from outside of the VOI (Fig. 5a). The phase of those signals at 1.5 ppm was variable, depending on the distance between




Fig. 2. Frequency and phase variation of the NAA signal in a data set measured in single-scan mode from the human brain at 7 T (NT = 160). **a** Uncorrected data; **b** data after automatic frequency and phase correction.




Fig. 3. In vivo ^1H NMR spectra of human brain (occipital gray matter) measured at 7 T using short-echo-time STEAM sequence with VAPOR water suppression. **a** Single-scan spectrum; **b** all 160 spectra from the study are superimposed. TE = 6 ms, TM = 32 ms, TR = 5 s, VOI = 8 ml. Processing: frequency and phase correction, eddy current correction (ECC), Gaussian multiplication $f(t) = \exp(-(t - t_{\text{max}})^2/2\sigma^2)$, $t_{\text{max}} = 0.0$ s, $\sigma = 0.085$ s, and fast Fourier transform (FFT).




Fig. 4. In vivo ^1H NMR spectrum measured from the human brain at 7 T. Sagittal and transverse RARE images illustrate the location of the VOI in the occipital lobe. Localization sequence: STEAM combined with OVS, VAPOR water suppression (Fig. 1), TE = 6 ms, TM = 32 ms, TR = 5 s, VOI = 8 ml, NT = 160. Processing: frequency and phase correction of single-scan FIDs, summation, ECC, Gaussian multiplication ($t_{\text{max}} = 0.05$ s, $\sigma = 0.085$ s), FFT, and zero-order phase correction.

the VOI and lipid-containing tissue. Water suppression without OVS was not as efficient and usually resulted in baseline deformation near the water resonance. It was difficult to completely eliminate eddy current effects resulting from gradient switching, in spite of dedicated efforts using quantitative eddy current field mapping [27]. Experimentally measured spectra without postprocessing ECC exhibited modest line shape and baseline distortions at these extremely short echo times (Fig. 5b).

3.7 Shimming

In addition to the localization performance of the sequence, shimming was extremely important to achieve a high spectral quality. The importance of the sec-




Fig. 5. Effects of localization performance, ECC, and shimming on the spectral quality. STEAM, TE = 6 ms, TM = 32 ms, TR = 5 s, VOI = 8 ml, NT = 32, $B_0 = 7$ T, human brainoccipital lobe. **a** OVS was turned off; **b** postprocessing ECC was not performed; **c** second-order shims were not used, linear shims were adjusted by FASTMAP; **d** reference spectrum (first- and second-order shims adjusted by FASTMAP, localization with STEAM and OVS, ECC performed). Gaussian multiplication ($\sigma = 0.1$ s) was applied.

ond-order shimming is demonstrated in Fig. 5c and d. Despite fine adjustment of first-order shims by FASTMAP, substantially increased line width relative to the reference spectrum was observed, when second-order shims were not used (Fig. 5c). Fully automatic B_0 field mapping and adjustment of all first- and second-order shims (X , Y , Z , XZ , YZ , Z^2 , $2XY$, $X^2 - Y^2$) by FASTMAP in combination with a strong second-order shim system (shim coils and shim drivers) provided high spectral resolution at a given field strength and particular brain location. This was confirmed by the observation (Fig. 6) that the signal line widths of metabolites were mostly independent of the size of the measured VOI (0.5–8.0 ml). Average demands for second-order shims in the occipital lobe at 7 T did not exceed 30% of their maximum strength. Performance of FASTMAP shim-




Fig. 6. Effect of the volume size on the spectral resolution at 7 T. STEAM, TE = 6 ms, TM = 32 ms, TR = 5 s, human brain gray-matter-rich occipital lobe. Processing: frequency correction, summation, Gaussian multiplication ($t_{\max} = 0.0$ s, $\sigma = 0.085$ s), FFT, zero-order phase correction.




Fig. 9. In vivo ¹H NMR spectra measured from gray-matter-rich occipital lobe (a) and periventricular white-matter-rich tissue (b) of 10 subsequent subjects at 4 T. The stack plot illustrates the high reproducibility of achieved spectral quality. STEAM, TE = 4 ms, TM = 42 ms, TR = 4.5 s, VOI = 15 ml, NT = 160, half-volume RF coil. Glucose intravenous infusion was applied during each study ([Glc]_{plasma} ≈ 15 mM). First- and second-order shims, RF power, and water suppression parameters were adjusted and set fully automatically, which requires less than 2 min.




Fig. 10. In vivo ^1H NMR spectrum measured from the brain of conscious (nonsedated) pediatric subject at 4 T with a TEM volume RF coil. STEAM, TE = 5 ms, TM = 43 ms, TR = 4.5 s, VOI = 15 ml, NT = 64. Data averaging: single-scan mode. Processing: frequency correction, ECC, Gaussian multiplication ($t_{\text{max}} = 0.05$, $\sigma = 0.085$), FFT, zero-order phase correction. Transverse RARE image (inset) shows the position of the VOI.

3.9 Routine Application of Spectroscopy at High Field

Routine use of the aforementioned spectroscopic technique at 4 T is illustrated in Fig. 9. In vivo ^1H NMR spectra were taken from consecutive examinations of 10 different subjects (patients with diabetes and healthy controls) as part of a study focused on brain glucose quantification. The spectral quality was highly reproducible and operator-independent. The H-1 signal of α -D-glucose at 5.23 ppm shows the robustness of the VAPOR water suppression that suppressed an extremely strong water signal with high efficiency (Fig. 9).

Another example illustrates the measurement of white-matter-rich periventricular tissue of a nonsedated child with a volume coil at 4 T (Fig. 10). Single-scan data averaging and subsequent frequency and phase corrections eliminated effects of the random motion of the pediatric patient during an experiment, which resulted in distortionless spectra and high spectral resolution.

4 Discussion

The goal of in vivo NMR spectroscopy is a precise and accurate quantification of the highest possible number of tissue metabolites with a high spatial resolution in the shortest measurement time. Recent rapid technical developments, especially the construction of strong whole-body magnets, rendered in vivo NMR spectroscopy of humans at high magnetic field possible. However, NMR spectroscopy at high magnetic field brings about specific challenges that must be met

In spite of fine adjustment of gradient preemphasis, residual eddy current effect can cause line shape and/or baseline distortions (Fig. 5). Removal of residual eddy current effects using unsuppressed water signal further improved spectral quality because it eliminated not only small residual eddy current effects but it also eliminated the modulation of FIDs (antiphase satellites in spectra) caused by mechanical vibrations.

High spectral quality is a precondition for reliable metabolite quantification. The principle of LCModel analysis is illustrated in Fig. 7. According to our knowledge, spin simulation has not been used to generate LCModel basis set, although simulated prior knowledge was used for time-domain [34] and frequency-domain [35] data analysis. More than fifteen metabolites (termed "neurochemical profile") were reliably quantified (with CRLB of <20%) from in vivo ^1H NMR spectra of the human brain measured at 7 T (Fig. 8). For example, the increased chemical shift dispersion at 7 T provided improved discrimination between glutamate and glutamine (CRLB of <5% in this study) and enabled separate reliable quantification of each of these brain metabolites. Changes in the glutamate-to-glutamine ratio may indicate pathological processes and impaired glutamate–glutamine cycling between neurons and glia [28, 36]. Such a broad number of metabolites quantified noninvasively from a small volume in the brain is a hallmark of in vivo NMR spectroscopy at very high magnetic fields. Increased precision and reliability of metabolite quantification at high fields is likely not only to establish in vivo ^1H NMR spectroscopy as an investigative tool to study the human brain but also to enhance the breadth of diagnostic applications.

5 Conclusions

High-field in vivo ^1H NMR spectroscopy of the human brain is a powerful method for the noninvasive quantification of a number of brain metabolites. Currently it is possible to perform in vivo ^1H NMR spectroscopy of humans at 4 and 7 T routinely with an automatic adjustment of sequence parameters that guarantees high reproducibility and high spectral quality.

Acknowledgements

The work was supported by NIH grants P41RR08079, R01NS38672, R01NS35192, M01RR00400, WM Keck and March of Dimes Birth Defects Foundations.

References

1. Wüthrich K.: *J. Biomol. NMR* **27**, 13–39 (2003)
2. Riederer S.J.: *Radiology* **231**, 628–631 (2004)
3. Tkáč I., Andersen P., Adriany G., Merkle H., Uğurbil K., Gruetter R.: *Magn. Reson. Med.* **46**, 451–456 (2001)
4. Terpstra M., Uğurbil K., Gruetter R.: *Magn. Reson. Med.* **47**, 1009–1012 (2002)

5. Michaeli S., Garwood M., Zhu X.H., DelaBarre L., Andersen P., Adriany G., Merkle H., Uğurbil K., Chen W.: *Magn. Reson. Med.* **47**, 629–633 (2002)
6. Uğurbil K., Adriany G., Andersen P., Chen W., Garwood M., Gruetter R., Henry P.G., Kim S.G., Lieu H., Tkáč I., Vaughan T., Van De Moortele P.F., Yacoub E., Zhu X.H.: *Magn. Reson. Imaging* **21**, 1263–1281 (2003)
7. Bottomley P.A.: *Ann. N.Y. Acad. Sci.* **508**, 333–348 (1987)
8. Ordidge R.J., Mansfield P., Lohman J.A., Prime S.B.: *Ann. N.Y. Acad. Sci.* **508**, 376–385 (1987)
9. Frahm J., Bruhn H., Gyngell M.L., Merboldt K.D., Hancicke W., Sauter R.: *Magn. Reson. Med.* **9**, 79–93 (1989)
10. Tkáč I., Starčuk Z., Choi I.Y., Gruetter R.: *Magn. Reson. Med.* **41**, 649–656 (1999)
11. Tkáč I., Henry P.G., Andersen P., Keene C.D., Low W.C., Gruetter R.: *Magn. Reson. Med.* **52**, 478–484 (2004)
12. Tkáč I., Andersen P., Adriany G., Gruetter R., Uğurbil K. in: *Proceedings of the 7th Scientific Meeting of the International Society for Magnetic Resonance in Medicine*, May 22–28, 1999, Philadelphia, Pennsylvania, USA (Grist T.M., ed.), p. 1612. Philadelphia: ISMRM 1999.
13. Tkáč I., Andersen P., Adriany G., Merkle H., Uğurbil K., Gruetter R. in: *Proceedings of the 8th Scientific Meeting of the International Society for Magnetic Resonance in Medicine*, April 4–7, 2000, Denver, Colorado, USA (Moonen C.T., ed.), p. 367. Denver: ISMRM 2000.
14. Tkáč I., Gruetter R. in: *Proceedings of the 8th Scientific Meeting of the International Society for Magnetic Resonance in Medicine*, April 4–7, 2000, Denver, Colorado, USA (Moonen C.T., ed.), p. 1856. Denver: ISMRM 2000.
15. Tkáč I., Seaquist E.R., Gruetter R. in: *Proceedings of the 8th Scientific Meeting of the International Society for Magnetic Resonance in Medicine*, April 4–7, 2000, Denver, Colorado, USA (Moonen C.T., ed.), p. 1938. Denver: ISMRM 2000.
16. Seaquist E.R., Damberg G.S., Tkáč I., Gruetter R.: *Diabetes* **50**, 2203–2209 (2001)
17. Öz G., Tkáč I., Choi I.Y., Bjoraker K.J., Shapiro E.G., Charnas L.R., Gruetter R.: *Neurology* **64**, 434–441 (2005)
18. Gruetter R.: *Magn. Reson. Med.* **29**, 804–811 (1993)
19. Gruetter R., Tkáč I.: *Magn. Reson. Med.* **43**, 319–323 (2000)
20. Adriany G., Gruetter R.: *J. Magn. Reson.* **125**, 178–184 (1997)
21. Vaughan J.T., Garwood M., Collins C.M., Liu W., DelaBarre L., Adriany G., Andersen P., Merkle H., Goebel R., Smith M.B., Uğurbil K.: *Magn. Reson. Med.* **46**, 24–30 (2001)
22. Provencher S.W.: *Magn. Reson. Med.* **30**, 672–679 (1993)
23. Provencher S.W.: *NMR Biomed.* **14**, 260–264 (2001)
24. Pfeuffer J., Tkáč I., Provencher S.W., Gruetter R.: *J. Magn. Reson.* **141**, 104–120 (1999)
25. Govindaraju V., Young K., Maudsley A.A.: *NMR Biomed.* **13**, 129–153 (2000)
26. Klose U.: *Magn. Reson. Med.* **14**, 26–30 (1990)
27. Terpstra M., Andersen P.M., Gruetter R.: *J. Magn. Reson.* **131**, 139–143 (1998)
28. Tkáč I., Keene C.D., Pfeuffer J., Low W.C., Gruetter R.: *J. Neurosci. Res.* **66**, 891–898 (2001)
29. Tkáč I., Rao R., Georgieff M.K., Gruetter R.: *Magn. Reson. Med.* **50**, 24–32 (2003)
30. Garwood M., DelaBarre L.: *J. Magn. Reson.* **153**, 155–177 (2001)
31. Tannus A., Garwood M.: *NMR Biomed.* **10**, 423–434 (1997)
32. Adriany G., van de Moortele P.F., Wiesinger F., Moeller S., Strupp J.P., Andersen P., Snyder C., Zhang X., Chen W., Pruessmann K.P., Boesiger P., Vaughan J.T., Uğurbil K.: *Magn. Reson. Med.* **53**, 433–445 (2005)
33. Van de Moortele P.F., Pfeuffer J., Glover G.H., Uğurbil K., Hu X.: *Magn. Reson. Med.* **47**, 888–895 (2002)
34. Ratiney H., Coenradie Y., Cavassila S., van Ormondt D., Graveron-Demilly D.: *Magma* **16**, 284–296 (2004)
35. Soher B.J., Young K., Govindaraju V., Maudsley A.A.: *Magn. Reson. Med.* **40**, 822–831 (1998)
36. Daikhin Y., Yudkoff M.: *J. Nutr.* **130**, 1026S–1031S (2000)

Authors' address: Ivan Tkáč, Center for Magnetic Resonance Research, Department of Radiology, University of Minnesota, 2021 6th Street SE, Minneapolis, MN 55455, USA
E-mail: ivan@cmrr.umn.edu