

Personal Mobility with Synchronous Trunk-Knee Passive Exoskeleton: Optimizing Human-Robot Energy Transfer

Diego Paez-Granados^{1†}, *Member, IEEE*, Hideki Kadone², Modar Hassan³, Yang Chen³, and Kenji Suzuki⁴, *Member, IEEE*

Abstract—We present a personal mobility device for lower-body impaired users through a light-weighted exoskeleton on wheels. On its core, a novel passive exoskeleton provides postural transition leveraging natural body postures with support to the trunk on sit-to-stand and stand-to-sit (STS) transitions by a single gas spring as an energy storage unit. We propose a direction-dependent coupling of knees and hip joints through a double-pulley wire system, transferring energy from the torso motion towards balancing the moment load at the knee joint actuator. Herewith, the exoskeleton maximizes energy transfer and the naturalness of the user’s movement. We introduce an embodied user interface for hands-free navigation through a torso pressure sensing with minimal trunk rotations, resulting on average $19^\circ \pm 13^\circ$ on six unimpaired users. We evaluated the design for STS assistance on 11 unimpaired users observing motions and muscle activity during the transitions. Results comparing assisted and unassisted STS transitions validated a significant reduction (up to 68% $p < 0.01$) at the involved muscle groups. Moreover, we showed it feasible through natural torso leaning movements of $+12^\circ \pm 6.5^\circ$ and $-13.7^\circ \pm 6.1^\circ$ for standing and sitting, respectively. Passive postural transition assistance warrants further work on increasing its applicability and broadening the user population.

Index Terms—Passive exoskeleton, Human-Robot Interaction, Standing Mobility Vehicle, Design Optimization

I. INTRODUCTION

POSTURAL transitions from sit-to-stand and stand-to-sit (STS) are common activities of daily living (ADL) usually overlooked by unimpaired people. However, any impairment or dysfunction on the lower-body results in hindering this crucial activity which initiates locomotion and allows social

¹ D. Paez-Granados is with the Swiss Federal Institute of Technology in Lausanne, EPFL, and is a Visiting Researcher at the Artificial Intelligence Laboratory, University of Tsukuba, Japan. dfpg@ieee.org

² H. Kadone is with the Center for Innovative Medicine and Engineering, University of Tsukuba Hospital, Japan. kadone@md.tsukuba.ac.jp

³ M. Hassan and Y. Chen are with the Artificial Intelligence Laboratory, Department of Intelligent Interaction Technologies, University of Tsukuba, Japan. hassan.chenyang@ai.iit.tsukuba.ac.jp

⁴ K. Suzuki is with the Faculty of Engineering and Center for Cybernetics Research, University of Tsukuba, Japan. kenji@ieee.org

Manuscript received May 31, 2021; revised November 2, 2021; accepted November 23, 2021. Recommended by Technical Editor J. A. Schultz and Senior Editor X. Chen. This work was supported in part by the Ministry of Education, Culture, Sports, Science and Technology of Japan under Grant 2019R300-2 and in part by the Toyota Mobility Foundation under the Mobility Unlimited Challenge grant. (Corresponding author: Diego Paez-Granados.)

This article has supplementary material provided by the authors and color versions of one or more figures available at <https://doi.org/10.1109/TMECH.2021.3135453>. Digital Object Identifier 10.1109/TMECH.2021.3135453




Fig. 1. Synchronous motion of knees joint and pelvis allow a lower-limbs and torso passive exoskeleton for supporting stand-to-sit and sit-to-stand transitions with a single passive element as external energy storage and controlled with natural motions of the upper-body.

and physical interaction with the surrounding environment and people. e.g., in cases of motor control dysfunctions, loss of muscle mass (muscular dystrophy), spinal cord injury, or amputation of limbs. All of which require full assistance to recover STS capability. Moreover, ageing by itself results in making sitting and standing transitions one of the most difficult and energy-expensive ADLs (high metabolic cost [1]) for most of the elderly life [2], [3]. Wheelchairs are a mobility solution, however, it completely removes standing from daily life which helps to preserve vital functions of the body, e.g., bone density, bowels function, cardiovascular and respiratory functions. Confirmed in studies of the consequence of sedentary behaviour [4], [5].

With this perspective, several assistive devices for STS and personal mobility are currently available such as power standing wheelchairs, and exoskeletons. Exoskeletons for lower-limb paralysis focus on assisting different mobility impairments, e.g., the robot suit HAL used in rehabilitation scenarios [6], the ReWalk exoskeleton for spinal-cord injury (SCI) persons, which relies on crutches for achieving locomotion [7]. While other research approaches envision soft robotic suits for hemiplegia patients [8] through a flexible exoskeleton for lower-limbs assistance with wire-driven systems [9], [10], [11] and other works as reviewed in [12].

Nonetheless, commercially available mobility devices are more common in the form of standing wheelchairs providing readily daily-life solutions that give more autonomy and speed

through motorized systems. However, one drawback is the significant power consumption in lifting the user and the heavier the devices become compared with seated powered wheelchairs e.g., TekRMD (Matia Robotics Inc. USA) standing wheelchair weights 120 kg with power autonomy of 5 hours. Therefore, end-users will be driven towards the more standard solution of a seated life.

Previous works have proposed lifting mechanisms for manual wheelchairs through hand-lever propelled systems as in [13], at the cost of removing standing locomotion. Our work aims to fill the gap of personal standing mobility through an energy-efficient daily-life mobility device compact and lightweighted, thus giving higher levels of autonomy to a broad range of users. We propose a mobile base capable of full walking speed and an energy-efficient STS passive exoskeleton with lower-limbs and trunk support by using natural upper-body postures, which is important for the acceptability and daily usage of such a device.

In previous work, we introduced PAL design [14] a device that assists ankles and knees joints through asymmetric transitions on a two stages STS transition based on static modelling of the human body, also extended to children's usage in [15]. Nonetheless, support to the torso during the transitions was not present, and it is required for users unable to hold straight postures (torso support) i.e., lack of control over rectus abdominals, erector spinae, or total lack of muscle control on lower-trunk as SCI of level T6 and above.

In the current work, we extended our previous design in [16] with a new system that supports the user's torso in both standing and sitting. We propose a novel double-coupling wire-driven parallel mechanism that isolates the sit-to-stand from the stand to-sit synchronized motions. Herewith, allowing an asymmetric postural transition with torso support (see, Fig. 1). Moreover, we present the design for a personal mobility device controlled through a pressure sensing system embedded on the exoskeleton at the abdominal area of the user for hands-free navigation, controlled through the algorithm we introduced in [17].

Compared with previously proposed devices for standing locomotion [14], [15], [16], our device coupled pulley system limits the range of users because of the non-linear multi-limb load to the passive element, tested between 55-72 kg (2-gas-springs) and 73-88 kg (3-gas-springs). Whereas the proposed device in [14] supported users from 60 - 100 kg (albeit without torso support). In comparison, standing power wheelchairs such as the LEVO (LEVO Ag, Switzerland) and Tek RMD (Matia Robotics Inc., USA) can lift users up to 120 kg. Nonetheless, our prototype of a passive exoskeleton with active wheels was achieved in a 36 kg compact system, whereas the LEVO is about 170 kg, and Tek RMD is 118 kg device.

The contributions of this paper are mainly two-fold: first, a novel design of a personal mobility device (PMD) for standing locomotion. Consisting of a passive exoskeleton for lower limbs and trunk support on a mobile-powered base controlled by torso motions in a hands-free control system. The novel mechanism is described in detail through a multi-objective optimal design methodology for personalising the human-




Fig. 2. Proposed simplified human-body model on the Saggital plane during STS transitions for analysis and design of the exoskeleton.

machine interface modelled through an impedance framework. Second, we demonstrated the performance of the device through a user study with unimpaired participants analyzing the STS assistance provided by the exoskeleton and compared it with the natural motions of the user.

II. METHOD

In this section, we describe the assumptions, objectives, and constraints used for investigating the models of interaction that could achieve the desired assistance in the sit-to-stand/stand-to-sit (STS) postural transitions.

In the following subsections, we first describe the dynamic model of the human body in the STS and propose the asymmetric postural transition that gives a user control over the motion by natural upper-body postures. Subsequently, we formulated the desired assistance for STS postural transition by synchronising the desired motions of the 2 main joints to be supported (namely knees and hip). Finally, we modeled the mechanism's kinematics and dynamics of the coupled motion with the user to a mathematical description in the context of human-robot interaction including the contact dynamics for simulating the STS transitions herewith, we constructed an optimization framework for evaluating multiple user-body characteristics in the design of the exoskeleton system.

A. Task Dynamics: Human Biomechanical Model

Previous works have focused on the joint torque support for STS transitions with 1 DOF torso, minimizing a jerk criterion on a 3-DOF powered system [18]; or through optimal trajectories of the COM [19], demonstrated with elderly patients in [20]. In contrast to previous works where the body model was reduced to 1 degree of freedom (DOF) at the torso [14], [16], [18], [21] we concluded that a 7-DOF human biomechanical model (HBM) was more adequate to fit the transitions shown in Fig. 2. Among the target users with spinal cord injuries (SCI) at thoracic nerves T6 and below, cerebral palsy (C.P.) or stroke, many upper-body functions remain feasible and could




Fig. 3. Postural transition support from sit-to-stand. State 1: the seated natural state where a user can freely move their torso, neither pulley system is engaged (P_N). State 2, sit-to-stand transition engaging through a torso's forward inclination (γ) enabling the first pulley system (P_1) in blue for torso support. State 3, standing. State 4, transition stand-to-sit with through a leaning backward (β) required for engaging the second pulley system (P_2). Finally, state 5 is the sitting posture showing the final torso angle (δ). Afterwards the cycle restart to state 1.

contribute to the postural transition, therefore, in our work we include crouching of the spine (q_4), arms (q_5, q_6, q_7) and head in the modelling in order to determine possible postures for controlling the device motion.

We modelled the dynamic load of the user to an exoskeleton supporting this transition through a Lagrangian formulation for an open kinematic chain [22], as follows:

$$\tau_{hm} = M_{hm}(q_{hm})\ddot{q}_{hm} + \eta_{hm}(\dot{q}_{hm}, q_{hm}), \quad (1)$$

where $\tau_{hm} = [\tau_1, \tau_2, \tau_3, \tau_4, \tau_5, \tau_6, \tau_7]^T$ represent the torques in the HBM joint space [ankle, knees, hip, torso, shoulder, elbow, wrist] as $q_{hm} = [q_1, q_2, q_3, q_4, q_5, q_6, q_7]^T$, as in Fig. 2(a). With a positive definite inertia matrix M_{hm} , and the term $\eta_{hm} = C(\dot{q}_{hm}, q_{hm})\dot{q}_{hm} + g(q_{hm})$ accounts for Coriolis-centrifugal, and gravitational forces respectively.

The center of mass (COM) for each segment can be estimated by a statistically equivalent serial chain (SESC) as described in [23],

$$C_1^M = \frac{m_1 A_0^1 \{c_1\}}{M} + \frac{m_2 A_0^2 \{c_2\}}{M} + \dots + \frac{m_n A_0^n \{c_n\}}{M}, \quad (2)$$

where A_0^i represents an homogeneous transformation matrix from the coordinate system 0 to i , with terms c_i noting the location of the center of mass m_i of the link i w.r.t. the coordinate system ($i - 1$), as depicted in Fig. 2(a). And the term M denotes the total body mass as $M = \sum_0^n m_i$, for $i \in R^n$.

We recorded the natural STS postural transitions of 7 unimpaired participants showing in Fig. 2(b) and extracted the loads and synchronous motion of the knee and hip using the proposed model. The main takeaway from these data was a common leaning forward method in the sit-to-stand transition where the hip angle - q_3 - reached angles between 30 to 45 degrees; validating a natural method of postural transitions through leaning the upper body for minimizing the knee joint torque.

B. Synchronous Asymmetric Postural Transitions

In this work, we proposed to achieve the STS transitions on asymmetric standing (leaning forward) and sitting motions

(leaning backward), which resulted from targeting the following objectives:

- 1) Standing and sitting support for lower-limbs and trunk.
- 2) Natural transition: controlled by the upper-body motion.
- 3) Synchronous motion of lower-body and upper-body.
- 4) Comfortable transitions: minimal jerk and max fit of actuator profile to load.
- 5) Low-weighted device: smallest possible passive element at a single location.

Figure 3 depicts the result of our design for passive transitions.

Stage 1, gives seated freedom to the user for moving his/her torso angles $\gamma + \delta$, with a safety lock on the exoskeleton to avoid undesired motions. **Stage 2**, starts by releasing the mechanical lock on the exoskeleton, followed by a leaning forward of the torso of an angle γ driven by the user. Afterwards, the movement of the knee joint should be coupled to the hips (represented by the blue pulley) making the exoskeleton's motion synchronous during the transition. **Stage 3** at the standing posture, the exoskeleton reaches a mechanical limit and the motion is halted. The user has leaning backwards freedom β , and the same mechanical lock to avoid undesired transitions from the exoskeleton. **Stage 4**, the user unlocks the exoskeleton and leans backwards engaging at an angle β where the system couples the hips to knee joint motions (depicted in red in Fig. 3), herewith giving support to the back of the user. **Stage 5**, this is the final state where the user arrives at sitting with an angle δ on their torso.

C. Coupling-Decoupling Wire-driven Parallel Joint

Preliminary testing with spinal cord injury volunteers on our previous prototype [16] showed that several users required support at their torso on both standing and sitting transitions because of the lack of control over their abdominal and back muscles. This varies significantly by the level of injury from T10 to T6 and each patient unique remaining muscle control.

Therefore, we propose a double wire-driven parallel mechanism to provide torso support in both transitions through the asymmetric motion where hips to knee coupled motions of sitting and standing differ in the expected coupled angles at the hips [$\gamma \sim -\pi/2$] for standing, and [$-\pi/2 + \beta \sim \delta$] for sitting. We found wire-driven transmissions a light-weighted solution to transfer power and couple multi-limb motions, as used in multi-joint serial robots, such as the WAM (Barret Technology, USA) and some novel joint transmission systems [24]. Here, the proposed design creates a controlled coupling-decoupling motion of two joints, making it directional dependent through a set of two different pulley circuits with independent length constraints that achieve the desired coupling angles, as illustrated in Fig. 4. Each transmission system engages separately for standing and sitting through the blue wire-pulley system P_1 and the red system P_2 , respectively.

The first pulley system P_1 (marked on a blue path for standing torso support in Fig. 4 left-side) operates by anchoring at a pulley on the hips joint q_3 (radius r_1) through a wire circuit exiting link 2 at w and connecting to the base link 1 at p , the wire exits link 1 at o , and anchors at v on link 2. The change in distance $\Delta(|p-w| + |v-o|)$ between sitting ($q_2 = q_o$) and




Fig. 4. Exoskeleton's optimization variables: we propose a double wire pulley mechanism that couples the motions at q_2 (knees) and q_3 (hips). The blue pulley system P_1 for sit-to-stand (left-side), couples the hips joint q_3 to q_2 , and controls the torque transfer through $|p-w|$, and $|v-o|$. Transferring load from link L_2 (A_2) to the base link L_1 (A_1). Correspondingly, the red pulley system P_2 (right-side), drives q_3 coupled to q_2 by connecting $|n-u|$, for the stand-to-sit transition. The whole COM expected movement in both transitions for the PMD w/ user and the user alone is depicted on the right side.

standing ($q_2 = q_f$) determines the output motion at q_3 , from $\gamma \sim q_s$. This closed pulley circuit allows to partially balance the moment non-linearity at q_2 caused by having the coupled motion of both joints.

Then, the leaning forward angle (γ) for engaging the first wire-pulley allows the motion to be driven by:

$$q_3|_{\gamma}^{q_s} = \alpha_1 q_2|_{q_o}^{q_f}, \quad (3)$$

where α_1 represents a linear gain, q_s denotes the final hip joint in standing, q_o and q_f represent the knee angle at sitting and standing.

Similarly, the second pulley system P_2 (marked on a red path for sitting torso support in Fig. 4 right-side) couples the motion through a pulley at q_3 (radius r_2) and the wire exiting link 2 at u , and fixing at n on link 1. Thus, the distance change $\Delta|n-u|$ between standing ($q_2 = q_f$) to sitting ($q_2 = q_o$) determines the torso motion q_3 , from $\beta \sim \delta$.

Then, the leaning backwards an angle (β) leads to the engaging the second mechanism whereas the first wire-pulley system remains out of range, thus the motion is driven by a second system:

$$q_3|_{\beta}^{\delta} = \alpha_2 q_2|_{q_f}^{q_o}, \quad (4)$$

where α_2 denotes an independent linear gain, and δ denotes the final hip joint at sitting ($q_2 = q_o$).

The location of these wire-pulley systems allows us to control the observed momentum at the knee joint, obtained as the sum of moments at q_2 :

$$M_o|_{q_s}^{q_f} = -\tau_2 + (v \times \bar{T}_i) + (w \times \bar{T}_o), \quad (5)$$

where v and w represent the location of the pulley system input and output on link 2. T_o denotes the force been transferred in the direction $|v-o|$ and T_i the output force in the direction $|p-w|$, as depicted in Fig. 5 right-side. The sitting wire-pulley system (depicted in red) transfers upper-body load (τ_3) to the knee joint through the connection in T_u (see, Fig. 5), thus, the load at the knee joint can be written as:

$$M_o|_{q_f}^{q_s} = -\tau_2 + (u \times \bar{T}_u), \quad (6)$$




Fig. 5. Left side: Abstraction of the exoskeleton interaction with a user through an impedance model. The user load F_h , represents the human as M_h , K_h , and D_h interfacing with a coupling K_i , D_i which accounts for the mechanical compliance of the structure and link. The exoskeleton M_r is controlled by an actuator F_a (resultant of K_a , D_a). Right side: free body diagram of the moment loads at the knee joint q_2 , showing the standing pulley system forces (blue), and sitting pulley system forces (red).

where u represents the location of the output pulley on link 2, and T_u the output force in the direction $|n-u|$ (shown in Fig. 4).

D. Human-Robot Interaction Multi-objective Optimization

The objectives proposed in section II-B have competing trade-offs of linearity, power transfer and user motion requirement through the proposed double-wire mechanism. Therefore, we modeled the STS transition considering the dynamics of the interaction and created a framework for simulating towards an optimal solution for both sitting and standing assistance mechanisms.

We set the features in $\xi = [u, v, w, n, o, p, r_1, r_2, \eta]$ as minimization arguments accounting for transmission through the standing wire pulley systems P_1 (v, o, p, w, r_1), the sitting wire pulley systems P_2 (n, u, r_2), and the efficiency of the transmission η . We optimize through the location of the vectors $|p-w|$, $|v-o|$ and the force-torque transferring ratio at the hip joint r_1 for standing. While $|u-n|$ and the transmission ratio r_2 , for sitting. **Constraints:** Location constraints were applied to achieve a feasible solutions around the user's body limiting the size of the exoskeleton. $(u, v, w) \in \mathbb{R}^2$ were constrained to an area A_2 ($u \wedge v \wedge w \in A_2$). $(n, o, p) \in \mathbb{R}^2$ correspond to the base link l_1 constrained to A_1 ($n \wedge o \wedge p \in A_1$).

As well, an equality constraint was set to enforce the start to end synchronous motion between of q_2 and q_3 , for sitting and standing through:

$$\Delta(|p-w| - |v-o|)|_{q_o}^{q_f} = r_1 \Delta q_{3std}|_{\gamma}^{q_s}, \quad (7)$$

$$\Delta(|n-u|)|_{q_f}^{q_o} = r_2 \Delta q_{3sit}|_{\beta}^{\delta}, \quad (8)$$

where r_1 represents the radius of the output pulley at the hip joint q_3 , and $\Delta q_{3std} = \gamma - q_s$ denotes difference between engaging angle at the leaning forward motion γ and the final standing posture angle q_s for standing transition, while $\Delta q_{3sit} = \beta - \delta$ denotes sitting transition difference from the leaning backward angle β and the sitting pose δ .

The optimization framework for achieving the best solutions within the space of feasible arrangements for the exoskeleton was evaluated by a Pareto-optimal front explored through a many-objective non-dominated sorting genetic algorithm

(NSGA-II) [25]. An exploration algorithm validated in convergence towards the actual optimal Pareto front in design examples, even under multiple state scenarios [26]. This algorithm allowed us to mutate over the possible populations of solutions through a definition of the objectives and associated costs for achieving the proposed mechanism, evaluated through the forward dynamics of the system as modelled in the previous section.

We model the interface through an impedance interaction at the two joints in contact with a damped interface, as it has been applied for stable physical human-robot interaction in whole body impedance control [27], and introduced in our previous work [16].

$$\tau_{hr} = M_h \ddot{q}_c + D_h \dot{q}_c + K_h q_c + \eta_i(\dot{q}_c, q_c), \quad (9)$$

where the η_i represents the non-linear terms defined in (1) for joint $i \in R^n$, considered for the serial kinematic chain. Nonetheless, for a user with paralyzed limbs the stiffness K_h and damping D_h of each joint are insufficient for holding the upright posture, thus, compared with the exoskeleton's it could be neglected yielding: $\tau_{hr} = M_h \ddot{q}_c + \eta_i(\dot{q}_c, q_c)$. Herewith, the equation of motion for the exoskeleton with a torque input from the actuator to the human body can be expressed as:

$$\tau_{hr} = (M_h + M_r) \ddot{q} + D_i \dot{q} + K_i q + \eta_i(\dot{q}, q), \quad (10)$$

where $\tau_2 = \tau_a$. And the effective torque at the joint is defined by the location of the linear actuator a and b , as depicted in Fig. 4. Thus, $\tau_a = \bar{a} \times \bar{F}_a$, where the actuator's force is defined ideally by the linear equation:

$$F_a = (f_0 + k_a \Delta x) \eta_t + D_a \dot{x}, \quad (11)$$

where D_a , and k_a denote the spring-damper characteristics of the actuator, f_0 represents the zero crossing force, and η_t represents the whole power transfer efficiency of the wire-driven system.

Using the above model for the moment load (10), transmission system (5 and 6) and actuator profile (11), we introduce the following definitions of the objectives for optimization:

1) *Objective 1 - Minimal moment load at the knee joint:* We defined the overall effective moment load (M_o) at the knee joint - q_2 in terms of the design parameters encapsulating the locations of each connection point between the thighs link (L_2) and the base (L_1), for both standing and sitting transitions as:

$$\begin{aligned} \operatorname{argmin}_{\xi} \quad & \frac{1}{M_r m} \sum_{q_2=q_o}^{q_f} M_o(\xi, \tau_{hr}, q_2), \\ \operatorname{argmin}_{\xi} \quad & \frac{M_r}{m} \sum_{q_2=q_f}^{q_o} \frac{1}{M_o(\xi, \tau_{hr}, q_2)}, \end{aligned} \quad (12)$$

where q_o denotes the sitting angle and q_f the standing angle at the knee joint, M_o represents the moment load at the knee joint of the exoskeleton, M_r corresponds to a normalization constant (max load for a 90 kg user), and m corresponds to the number of samples used in the discrete space of $q_2 = [q_o \sim q_f]$, as $m = (q_f - q_o)/dq$ for a sample rate dq .




Fig. 6. Pareto-optimal front of possible configurations to achieve the synchronous transitions with the proposed objectives. The black dots represent the discrete solutions for each objective: moment load at q_2 (12), torque linearity at q_2 (14), and motion linearity between knees and hip (13). The surrounding contour was derived by scatter interpolation. The chosen solution was set at the marked red star, where we chose based on the linearity of the motion and a mid-point to other objectives.

2) *Objective 2: Natural torso motion:* We minimized the difference to the natural posture transition based on the captured data presented in Fig. 2(a), which we approximated as a linear motion of the knees (q_2) and hips (q_3) within the range $q_3 = [\gamma \sim q_s]$ for standing, and $q_3 = [\beta \sim \delta]$ for sitting. Therefore, we minimize a linear fitting as follows:

$$\begin{aligned} \operatorname{argmin}_{\xi} \quad & \frac{1}{(\gamma - q_s)m} \sum_{q_2=q_o}^{q_f} \|q_3(\xi, q_2) - q_l(\gamma, q_s, q_2)\|, \\ \operatorname{argmin}_{\xi} \quad & \frac{1}{(q_s - \beta)m} \sum_{q_2=q_f}^{q_o} \|q_3(\xi, q_2) - q_l(\beta, \delta, q_2)\|, \end{aligned} \quad (13)$$

where q_3 denotes the coupled hip joint during the transitions, q_s the final standing angle.

3) *Objective 3: Matching a linear torque profile:* We define both standing and sitting as,

$$\begin{aligned} \operatorname{argmin}_{\xi, q_{hr}} \quad & \frac{1}{M_n m} \sum_{q_2=q_s}^{q_f} \|M_o(\xi, \tau_{hm}, q_2) - M_l(q_o, q_f)\|, \\ \operatorname{argmin}_{\xi, q_{hr}} \quad & \frac{1}{M_n m} \sum_{q_2=q_f}^{q_s} \|M_o(\xi, \tau_{hm}, q_2) - M_l(q_f, q_o)\|, \end{aligned} \quad (14)$$

where $M_l(q_o, q_f)$ denotes the ideal moment at the knee joint for a passive linear actuator (the gas-spring) with force equation in (11).

The results in Fig. 6 present a set of contours that map the Pareto front with black circles for the three objectives. The trade-off on objectives 1, 2, and 3 exists as a result of the task dynamics and parallel mechanism, where the motion control could be achieved by either reducing significantly the moment load at the knee joint or the motion linearity, and the moment linearity is affected by the distribution of the pulley system. If a single objective were optimized, e.g., the moment load (as in our previous prototype [16]) the linearity of the hip-to-knees joint motion would be drastically reduced in the transition.

4) *Minimal Actuator Force:* Finally, the location of the passive actuator is independently optimized to maximize the torque output:

$$\operatorname{argmax}_{\Phi} \quad \sum_{q_2=q_o}^{q_f} \tau_a(\Phi), \quad (15)$$

where $\Phi = [F_a, \bar{a}, \bar{b}, n] \bar{b}a$ was chosen so that the actuators effective force F_a would satisfy the required torque to allow both sitting and standing transitions. Therefore the constraints were set for standing load $M_o(q) \big|_{q_f=q_o}$ and sitting load $M_o(q)_{q_2=q_f, q_o}$ as,

$$\tau_a > M_o(\xi, q_{hr}) \big|_{q_f=q_o}, \quad (16)$$

$$\tau_a < M_o(\xi, q_{hr}) \big|_{q_o=q_f} \quad (17)$$

Moreover, the actuator was selected from existing vendors database, thus, discretely selected to fit in the constrained area A_2 and A_1 .

E. User Embodied Control Interface

The exoskeleton system was designed as a personal mobility device (PMD) with a user intention recognition embedded on the torso support bar (link 3) through pressure sensing system. We propose a hands-free navigation method with small upper body motions by continuous mapping of the user's torso to the mobile exoskeleton's local coordinate system at the wheels (as shown in Fig. 7) by following the principle of gaze tracking during walking [28], which showed that a person would normally follow their walking direction with their head and torso; matching a natural walking motion while exploiting the user's upper-body residual capabilities. We mapped the upper-torso motion counter-clockwise (CCW) to a rotation CCW, a spinning clockwise (CW) to a CW rotation, and slight forward leaning of the torso to a forward driving at the base of the robot. The backward motion was defined by a conscious independent action of the hands with the use of the sensors at each extreme of the torso bar by pressing them simultaneously, thus, avoiding unintended activation.

As depicted in Fig. 7, we constructed the inner surface of the support bar with a soft material and an array of pressure sensors for mapping the body postures to a locomotion intention. The intention recognition algorithm maps the 2D projection of the pressure distribution changes to the robot's control space in velocity $\zeta = [v, \omega]'$, where v denotes the linear, and ω the angular velocity, both in the local frame of the robot. We defined the input space for the user $u = [\rho, P]$ as the center of pressure (ρ) and the maximum pressure (P) over the sensors' array. Subsequently, we used a proportional control to drive the robot $\|\zeta\|$ with the user given pressure P , $\|\zeta\| = KP$, with a gain matrix $K = [k_1, k_2]'$, constrained to the velocity limits of the mobility device (v_{max}, ω_{max}). A detailed description of the sensor mapping $v(\rho), \omega(\rho)$ and its generalization used in this controller can be found in our previous work [17].

III. SYSTEM CONFIGURATION: PERSONAL STANDING MOBILITY

Results from the selected configuration of pulley systems and actuators are shown in Fig. 8. We chose a value with higher priority to the motion linearity with the compactness of the location of the wire-driven system. One of the lowest moment load to linearity configurations was chosen (the red mark in Fig. 6). The simulation of multiple users' STS transitions with the user load to the knees in yellow (M_o), and the





Fig. 7. User control interface by a pressure sensing of upper-body motions which maps natural body postures to low-level motion control of the in-wheel motor system for driving mobility device.


exoskeleton with user transfer power to the knee joint plotted in blue (τ_2). The motion from the sitting posture starts on the top of the figure with loads over 200 Nm, which are controlled by the user through leaning backwards or forward. If the user-controlled motion overcomes the actuator generated torque in a seated posture (plotted as a grey line) at the "γ" zone the standing transition starts with the 1st pulley system engaging at q_{3std} (which varies with the user weight distribution). When reaching the standing zone (below $q_3 = -90^\circ$) the user has leaning backwards freedom but no leaning forward, as the engaged 1st pulley system reaches a mechanical limit through knees-joint at a set $q_2 = 90^\circ$. Around the leaning backwards "β" zone, the user-initiated motion increases the load making it larger than the provided by the actuator in standing (grey dotted line for the Sitting transition), therefore, a stand-to-sit transition motion starts which engages the 2nd pulley system that ensures support to the torso during sitting, so that, the user hip-joint reaches an angle q_{3sit} .

A. Exoskeleton Setup with Passive Actuator


The location of a gas-spring actuator was chosen by the system through objective 4, which was fit with multiple providers of commercial actuators. In this case, we chose not to fabricate a custom-made actuator rather, minimize costs and feasibility of applicability of the method by using a discrete number of off-the-shelf actuators. The significant hysteresis on gas-spring actuators was reported in [14], thus in Fig. 8(b) the red-line marks an ideal actuator whereas real torque outputs from data of the gas spring were depicted as a grey line for sit-to-stand and as a dotted grey line for stand-to-sit. From an ideal actuator equation in (11) provided by the manufacturer the data of actuation torque showed less than 80% of effective torque output whereas fitting an equation of measured forces to a third order polynomial showed 99% confidence as $F_a = (\lambda_0 + \lambda_1 \delta x + \lambda_2 * \Delta x^2) \eta + D_a \Delta x$. Therefore, the optimization process followed the fitted equation with variance in compression and extension forces to achieve an optimal placement.



(a) Moment load of the user at the knee joint (q_2) compared to the knee standing joint angle.



(b) Moment load at the knee joint (τ_2) during the stages of STS as depicted in Fig. 3 for multiple user's weight and height.




(c) COM motion required for operating the passive exoskeleton during STS transitions for users 42 – 62 kg with height range 1.40 – 1.80 cm.


Fig. 8. Pareto-optimal implementation and evaluation of potential user usability gap through the forward dynamic model and estimated weight/height distributions.

Using the selected setup we present the results of the model evaluation for the system usability of multiple user-body sizes with variations on height, weight and inertial distribution based on models of inertia distributions in [29], implemented through the forward dynamics in (10).

Fig. 8(c) shows the COM motion during the full STS, showing only users with full feasible STS in the current configuration. The data exemplify the required small movement by the user of his/her COM of 5 to 6 cm on the horizontal axis for initiating the sit-to-stand transition passively assisted. Equally, for stand-to-sit transitions a small variation of 5 to 8 cm is sufficient to control the transition. Also, the combined human-robot couple variations of the COM with the wheeled



(a) Lateral view of the mobility device with a virtual user. The trunk support (link 3) in orange, thighs holding (link 2) in black, and the shank base (link 1) in ochre fixed to the mobile base.



(b) Detailed view of the parallel cable-driven double-pulley system for transmission of the single actuator torque to both joints.

Fig. 9. Synchronous passive exoskeleton with lower-limbs and torso support with a single passive gas-spring as external energy storage, implemented on a mobile base system for standing mobility.

based which was an important factor for design of the static stability of the base, and should be consider for the dynamic stability for setting the motion control acceleration constraints.

B. Embodied Standing Mobility Device

The implemented exoskeleton is compact with a 3D printed surfaces on the frontal body (top of the legs and torso) and flexible belts for the rear body part. The overall weight is 6.5kg for links 2 and 3, and 10kg for the base link (including three gas-spring actuators). In order to fit multiple users we developed the current prototype by adjustable mechanism and belts at shank, thighs, buttocks, and back supports, as depicted in Fig.9 (see the attached multimedia for further details). The distribution over link 1 and link 2 in Fig.9(a), shows the assembly result matching the Pareto-optimal result for the wire pulley systems (standing mechanism in blue and sitting in red) and actuator's location. The wire-pulley systems achieved with stainless steel wires (2mm and 3mm) duplicated symmetrically on each side of the exoskeleton.




Fig. 10. Control system architecture, considering the low-level velocity controller for the mobile base of the robot, and the higher-level user intention recognition coupling.

The double pulley system (Fig. 9(b)) was embedded on a joint q_3 , and travels within link 2 through routing friction-less tubes. We achieved a flexible configuration of the engaging angles (γ and β) by changing the length and tension on the wire systems (P_1 and P_2) through a set of turnbuckles located at link 1 between pulleys ($p-b$) and ($u-n$). For user safety and ensuring fixed postures, an important component is a locking system through a manual lever located at the thighs, which allows the user to stop the postural transition by blocking the gas-spring power, thus, fixing the exoskeleton at any posture. As well, we include a end-switch at the standing and sitting posture to ensure the safe locomotion only at standing posture, while any other posture would engage the mechanical brakes.

C. Motion Control System

In the current design, we embedded 10 pressure sensors over a soft band around the torso bar for a user's control interface following the description in section II-E. The current sensors measure 25×25 mm, without spacing, so that the total covered length is 250 mm (shown in Fig. 9(b)), which fits a wide range of user's waist. The system was designed to provide high-level intention recognition and results in a desired velocity command (depicted in Fig. 10). Thus, the user interface controller was mounted on an embedded computer UP-Squared (AAEON Technology Inc.) using a High-Precision AD/DA Board (Waveshare) for sensors input and communication through analog channels to the actuation low-level control. A second low-level controller on an embedded micro-controller circuit handles acceleration and velocity limits and communicating through CANBUS to the wheels velocity controller. Finally, motion commands are executed on a powered wheeled system using commercial in-wheel motors (JWX-02, YAMAHA Motor Corp., Iwata, Japan) with an internal velocity control loop executing at 400 Hz.

IV. SYSTEM EVALUATION

A. Sit-to-Stand / Stand-to-Sit Assistance Evaluation

We conducted experiments with with ten unimpaired male volunteers (age 26.1 ± 4.6 years, 170.3 ± 8.9 cm, 67.6 ± 8.1 kg), comparing the necessary muscle groups activity with and

without the device, as well as, motion and characteristics of the support. The experiments proceeded as follows: each participant performed sit-to-stand and stand-to-sit postural transitions (Fig. 11(a)) over three times using the device and then without it (see, Fig. 11(b)). The participants were equipped with wireless Electromyography (EMG) sensors (Trigno Lab, Delsys, USA) on the bilateral extensor muscles; latissimus dorsi (LD) and erector spinae (ES) for trunk extension evaluation, gluteus maximus (GM) for hip extension, and vastus medialis (VM) for knee extension. Markers of a motion capture system (MX System, Vicon Motion Systems, Ltd., U.K.) were attached on the Acromion, Great Trochanter, Lateral Epicondyle of Femur, and Lateral Malleolus to record sagittal motion synchronized with EMG recordings, in order to detect the start and end of transitions (see, Fig. 11(c)). EMG data were band-pass filtered (40-300Hz), rectified, and evaluated according to maximum value of local integration by a moving window of 100 ms width.

A paired t-test was adopted to compare between the two conditions. To evaluate the amount of reduction, reduction ratio was computed by $R_r = 1 - \frac{EMG_{assisted}}{EMG_{without}}$, where $EMG_{assisted}$ and $EMG_{without}$ indicate the EMG with and without the assistance of the device respectively.

In the assisted sit-to-stand case, maximum activation levels were significantly reduced in average by 18.2% for ES, 55.6% for GM and 68.4% for VM when compared to that of the unassisted case. In the assisted case of stand-to-sit, activation levels were reduced in average by 54.7%, 41.3% and 62.5% respectively (Fig. 11(f)). With all significant levels taken for ($p < 0.01$). On the other hand, the activation of LD showed non-significant increase by $99.0 \pm 168.4\%$ and $33.4 \pm 103.2\%$ in sit-to-stand and stand-to-sit transitions respectively.

Subsequent evaluation of a user (88kg, 1.75m) load to the exoskeleton (see, Fig.11(d)) was performed by analysis the acceleration at each joint with additional markers placed at T6, C7, elbow, wrists, and hands, and with measured actuator force output when using 3-gas-springs. We found a sit-to-stand load normalized moment of 0.026 (well below the design value of 0.86). This is explained because of a significantly high leaning posture in the transition chosen by the subject where leaning forward at q_4 and q_5 resulted in most of the load quickly converging to the negative plane. For the stand-to-sit transition, the normalized load was 0.78, below the expected 0.86. We also observe a different load profile to the modelled motion. The linearity of the output torque showed an absolute normalized error of 0.098 for the sit-to-stand and 0.039 on the stand-to-sit transition, both below the expected value of 0.45 from simulation data.

The naturalness of the motion (Eq. 14) was validated among all subjects comparing natural transitions and assisted one (see, Fig. 11(b)). The normalized linearity of the coupled motion between joints $q_2 - q_3$ on the sit-to-stand transition was 0.11 ± 0.08 , and 0.09 ± 0.02 for stand-to-sit, both below the simulated expected value (0.15). Moreover, comparing to the natural transition of each subject and found an average γ of $12.3^\circ \pm 6.5^\circ$, which reflects an average $2.6^\circ \pm 16.5^\circ$ increment from the natural motion. All below the expected values of $25 \sim 38^\circ$ (see, Fig. 11(g)). The stand-to-sit transitions required




Fig. 11. User evaluation study for the proposed synchronous passive exoskeleton with lower-limbs and torso support through a single passive gas-spring as external energy storage, implemented on a mobile base system for standing mobility.

an inclination angle β of $-13.7^\circ \pm 6.1^\circ$, ($-6.0^\circ \sim -21.7^\circ$), all below the expected values of $-24 \sim -37^\circ$ (see, Fig. 11(h)). The transition times were on average $7.13 \pm 2.7s$ for sit-to-stand, and $6.7 \pm 1.9s$ in stand-to-sit. Which were in average an increment from the natural transitions of $4.6s$ and $4.1s$, respectively.

B. Motion Control through Upper-Body Pressure Sensing

Six unimpaired subjects were recorded performing circular and 8-like navigation motions with the embodied interface of control, as depicted in Fig. 12 with pressure sensing normalized by the sensors range (approximately $0.16N/cm^2$), and the mapped mobility device velocity response $\xi = [v, \omega]'$. We

found a peak body rotation by comparing shoulder to hip angle on the top plane with a mean of $0.34rad \pm 0.24rad$ ($19^\circ \pm 13^\circ$) among the participants. As well, we observed the change in EMG activity at ES and rectus abdominal (RA) muscles during the circuits. Results showed decreased ES activity (36% on CCW and 32% CW) and increased RA activity (14%) for both rotations compared with straight motions. These results show the user resulting motion for navigation as proposed through upper-body rotations or stiffening of the abdominal area by means of abdominal contractions and back extensions.




Fig. 12. Data example of a user pose to directional control space transformation based on pressure distribution sensing at the lower-torso.

V. DISCUSSION AND CONCLUSION

The proposed light-weighted PMD (36 kg) has a comparable autonomy to that of power wheelchairs (5 to 7 hours) and provides standing locomotion with hands-free navigation through pressure sensing of torso rotations ($\pm 19^\circ$). The passive exoskeleton allows natural torso postures for user-controlled STS transitions within short times of 7.3s and 6.7s, respectively. Faster than any standing wheelchair (20 ~ 60s).

Compared with our previous designs for ankle-knees support [14] and preliminary torso support for sit-to-stand [16], the novel exoskeleton showed a statistically significant reduction of the lower back, hip and knees extensor muscles (ES, GM and VM) during both postural transitions in sitting and standing, compared to natural transitions. These results suggest the possibility of usage for patients with a significant reduction in the lower back and lower limb muscle control would be able to perform the postural transitions using passive assistance. However, higher levels of mobility impairment might not achieve the device control due to the need for upper back muscles (LD) which showed no reduction.

Results of the dynamic load to the exoskeleton during interaction with a user showed the designed large difference in standing and sitting loads (Fig. 11(d)) thanks to the asymmetric postural transition while displaying a highly linear load at the knee joint. Although we found a significantly steeper slope on the moment load for the evaluated user (observed by the crouching of the back and head inclination taken by the user), the transitions were smooth given the damped response on the actuation (D_a).

The chosen Pareto optimal configuration showed to perform

within the designed parameters in the linearity of the coupled motion of thighs and torso support, with the coupling angles varying per user as expected (see, Fig. 11(g)) due to the physiological differences in mass distribution over the body. Nonetheless, the results of the leaning forward ($12.3^\circ \pm 6.5^\circ$) and backwards angles ($-13.7^\circ \pm 6.1^\circ$) were lower than simulations for all users, which validates our human models as conservative for the optimization. The only posture that exceeded expected values was the sitting angle (δ), with an average of 66.2 ± 19.9^{circ} , over the expected value of 45^{circ} . This might be caused by the user's loose attachment to the back support which needs improved fit to the user's height.

Further applications of the proposed exoskeleton and passive assistance methodology through energy balance of the user's motion are envisioned in rehabilitation scenarios through changes in the passive actuation, and further improvement by adapting the system for a wider range of user body types, including women, the elderly and the over-weighted population.

ACKNOWLEDGMENT

We thank TESIS graphics for the illustrations and graphical abstract. All user studies were approved by the University of Tsukuba, approval No.2019R300-2.


REFERENCES

- [1] T. Nakagata, Y. Yamada, Y. Hatamoto, and H. Naito, "Energy expenditure of a single sit-to-stand movement with slow versus normal speed using the different frequency accumulation method," *Medicina (Lithuania)*, vol. 55, no. 3, 2019.
- [2] U. Lindemann, R. Muehe, M. Stuber, W. Zijlstra, K. Hauer, and C. Becker, "Coordination of strength exertion during the chair-rise movement in very old people," *Journals of Gerontology - Series A Biological Sciences and Medical Sciences*, vol. 62, no. 6, pp. 636–640, 2007.
- [3] L. Ferrucci, R. Cooper, M. Shardell, E. M. Simonsick, J. A. Schrack, and D. Kuh, "Age-related change in mobility: Perspectives from life course epidemiology and geroscience," *Journals of Gerontology - Series A Biological Sciences and Medical Sciences*, vol. 71, no. 9, pp. 1184–1194, 2016.
- [4] N. Owen, G. N. Healy, C. E. Matthews, and D. W. Dunstan, "Too much sitting: The population health science of sedentary behavior," *Exercise and Sport Sciences Reviews*, vol. 38, no. 3, pp. 105–113, 2010.
- [5] P. T. Katzmarzyk and I. M. Lee, "Sedentary behaviour and life expectancy in the USA: A cause-deleted life table analysis," *BMJ Open*, vol. 2, no. 4, pp. 1–8, 2012.
- [6] A. Tsukahara, Y. Hasegawa, K. Eguchi, and Y. Sankai, "Restoration of Gait for Spinal Cord Injury Patients Using HAL with Intention Estimator for Preferable Swing Speed," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 23, no. 2, pp. 308–318, 2015.
- [7] G. Zeilig, H. Weingarden, M. Zwecker, I. Dudkiewicz, A. Bloch, and A. Esquenazi, "Safety and tolerance of the ReWalk™ exoskeleton suit for ambulation by people with complete spinal cord injury: A pilot study," *The Journal of Spinal Cord Medicine*, vol. 35, no. 2, pp. 101–96, 2012.
- [8] L. N. Awad, J. Bae, K. O'Donnell, S. M. De Rossi, K. Hendron, L. H. Slood, P. Kudzia, S. Allen, K. G. Holt, T. D. Ellis, and C. J. Walsh, "A soft robotic exosuit improves walking in patients after stroke," *Science Translational Medicine*, vol. 9, no. 400, 2017.
- [9] Y. Lee, J. Lee, B. Choi, M. Lee, S. G. Roh, K. Kim, K. Seo, Y. J. Kim, and Y. Shim, "Flexible gait enhancing mechatronics system for lower limb assistance (GEMS L-Type)," *IEEE/ASME Transactions on Mechatronics*, vol. 24, no. 4, pp. 1520–1531, 2019.
- [10] L. Zhou, W. Chen, J. Wang, S. Bai, H. Yu, and Y. Zhang, "A Novel Precision Measuring Parallel Mechanism for the Closed-Loop Control of a Biologically Inspired Lower Limb Exoskeleton," *IEEE/ASME Transactions on Mechatronics*, vol. 23, no. 6, pp. 2693–2703, 2018.

- [11] T. Zhang, M. Tran, and H. Huang, "Design and experimental verification of hip exoskeleton with balance capacities for walking assistance," *IEEE/ASME Transactions on Mechatronics*, vol. 23, no. 1, pp. 274–285, 2018.
- [12] G. Bao, L. Pan, H. Fang, X. Wu, H. Yu, S. Cai, B. Yu, and Y. Wan, "Academic review and perspectives on robotic exoskeletons," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 27, no. 11, pp. 2294–2304, 2019.
- [13] J. Shaikh-Mohammed, S. S. Dash, V. Sarda, and S. Sujatha, "Design journey of an affordable manual standing wheelchair," *Disability and Rehabilitation: Assistive Technology*, vol. 0, no. 0, pp. 1–11, 2021. [Online]. Available: <https://doi.org/10.1080/17483107.2021.1892839>
- [14] Y. Eguchi, H. Kadone, and K. Suzuki, "Standing Mobility Device with Passive Lower Limb Exoskeleton for Upright Locomotion," *IEEE/ASME Transactions on Mechatronics*, vol. 23, no. 4, pp. 1608–1618, 2018.
- [15] K. Sasaki, D. Paez Granados, M. Sugimoto, T. Sugiyama, and K. Suzuki, "Child-Sized Passive Exoskeleton for Supporting Voluntary Sitting and Standing Motions," in *IEEE International Conference on Intelligent Robots and Systems*. Madrid: IEEE Xplore, 2018, pp. 5457–5462.
- [16] D. F. Paez-Granados, H. Kadone, and K. Suzuki, "Unpowered Lower-Body Exoskeleton with Torso Lifting Mechanism for Supporting Sit-to-Stand Transitions," in *IEEE International Conference on Intelligent Robots and Systems*. Madrid: IEEE Xplore, 2018, pp. 2755–2761.
- [17] Y. Chen, D. Paez-Granados, H. Kadone, and K. Suzuki, "Control Interface for Hands-free Navigation of Standing Mobility Vehicles based on Upper-Body Natural Movements," in *IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS-2020)*. Las Vegas, USA: IEEE Xplore, 2020.
- [18] A. Asker, S. F. Assal, M. Ding, J. Takamatsu, T. Ogasawara, and A. M. Mohamed, "Modeling of natural sit-to-stand movement based on minimum jerk criterion for natural-like assistance and rehabilitation," *Advanced Robotics*, vol. 31, no. 17, pp. 901–917, 2017. [Online]. Available: <https://doi.org/10.1080/01691864.2017.1372214>
- [19] M. Geravand, P. Z. Korondi, C. Werner, K. Hauer, and A. Peer, "Human sit-to-stand transfer modeling towards intuitive and biologically-inspired robot assistance," *Autonomous Robots*, vol. 41, no. 3, pp. 575–592, 2017.
- [20] C. Werner, M. Geravand, P. Z. Korondi, A. Peer, J. M. Bauer, and K. Hauer, "Evaluating the sit-to-stand transfer assistance from a smart walker in older adults with motor impairments," *Geriatrics and Gerontology International*, vol. 20, no. 4, pp. 312–316, 2020.
- [21] H. R. Yamasaki, H. Kambara, and Y. Koike, "Dynamic optimization of the sit-to-stand movement," *Journal of Applied Biomechanics*, vol. 27, no. 4, pp. 306–313, 2011.
- [22] R. Featherstone and D. E. Orin, "Dynamics," in *Handbook of Robotics*, 1st ed., B. Siciliano and K. Oussama, Eds. Springer, 2008, ch. 2, pp. 35–62.
- [23] S. Cotton, A. Murray, and P. Fraise, "FOCUSED SECTION SHORT PAPERS-Estimation of the Center of Mass: From Humanoid Robots to Human Beings," *IEEE ASME Transactions on Mechatronics*, vol. 14, no. 6, pp. 707–712, 2009.
- [24] S. Jiang, D. Hua, Y. Wang, F. Ju, L. Yin, and B. Chen, "Design and modeling of motion-decoupling mechanism for cable-driven joints," *Advances in Mechanical Engineering*, vol. 10, no. 5, pp. 1–10, 2018.
- [25] K. Deb, A. Pratap, S. Agarwal, and T. Meyarivan, "A fast and elitist multiobjective genetic algorithm: NSGA-II," *IEEE Transactions on Evolutionary Computation*, vol. 6, no. 2, pp. 182–197, 2002.
- [26] K. Deb, L. Zhu, and S. Kulkarni, "Handling Multiple Scenarios in Evolutionary Multiobjective Numerical Optimization," *IEEE Transactions on Evolutionary Computation*, vol. 22, no. 6, pp. 920–933, 2018.
- [27] D. F. Paez Granados, B. A. Yamamoto, H. Kamide, J. Kinugawa, and K. Kosuge, "Dance Teaching by a Robot : Combining Cognitive and Physical Human – Robot Interaction for Supporting the Skill Learning Process," *IEEE Robotics and Automation Letters*, vol. 2, no. 3, pp. 1452–1459, 2017.
- [28] M. Hollands, A. Patla, and J. Vickers, "'Look where you're going!': Gaze behaviour associated with maintaining and changing the direction of locomotion," *Experimental Brain Research*, vol. 143, no. 2, pp. 221–230, 2002.
- [29] P. D. Leva, "Adjustment to Zatsiorsky-Seluyanov's Segment Inertia Parameters," *Journal of Bio*, vol. 29, no. 9, pp. 1223–1230, 1996.



Diego F. Paez-Granados received the Ph.D. degree in bioengineering and robotics from Tohoku University, Sendai, Japan, in 2017. He is currently a Researcher with the Learning Algorithms and Systems Laboratory, EPFL, Switzerland, and a Visiting Researcher with the University of Tsukuba, Tsukuba, Japan. His research interests include physical and cognitive human modeling, compliance in human-robot interaction, shared-control for robot navigation, and soft-robot design and control with human-in-the-loop.



Hideki Kadone received his Ph.D. degree in information science and technology from the University of Tokyo, Japan, in 2008. He is an associate professor in Faculty of Medicine, and Center for Cybernetics Research, at University of Tsukuba. His research interests include clinical motion measurement and analysis and development of wearable assistive devices to compensate for, support or improve motor function of patients after neurological or mechanical disorders.



Modar Hassan received the BS degree in Engineering in 2009 from Tishreen University Syria. He received ME and PhD in Engineering in 2013, 2016, and MS in Medical Sciences in 2016 from the University of Tsukuba, Japan. From 2010 to 2011 he was a lecturer of Mechatronics at Tishreen University, Syria. He is currently an assistant professor at the Faculty of Engineering, Information and Systems, University of Tsukuba, Japan. His research interests include assistive robotics, orthotics, biomechanics, and motor control. He is a member of the IEEE.



Yang Chen received the B.Eng. in mechanical engineering from Jilin University, China, in 2017 and his master degree in human informatics from the University of Tsukuba, Japan, in 2020 where he is currently working toward the Ph.D. degree. His research interests include human assistive device, mobile robot and autonomous control.



Kenji Suzuki received his PhD degrees in pure and applied physics from Waseda University, Japan, in 2003. He is currently a full professor in the Center for Cybernetics Research, and principal investigator of Artificial Intelligence Laboratory, Faculty of Engineering, University of Tsukuba, Japan. He was also a visiting researcher at the Laboratory of Physiology of Perception and Action, College de France in Paris, and the Laboratory of Musical Information, University of Genoa, Italy. His research interests include augmented human technology, assistive and

social robotics, humanoid robotics social playware, and affective computing.